



2655 Park Center Dr., Suite A
Simi Valley, CA 93065
T: +1 805 526 7161
F: +1 805 526 7270
www.alsglobal.com

LABORATORY REPORT

January 16, 2016

Glenn La Fevers
Southern California Gas Company
12801 Tampa Ave
Northridge, CA 91326-1045

RE: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

Dear Glenn:

Enclosed are the results of the samples submitted to our laboratory on January 15, 2016. For your reference, these analyses have been assigned our service request number P1600187.

All analyses were performed according to our laboratory's NELAP and DoD-ELAP-approved quality assurance program. The test results meet requirements of the current NELAP and DoD-ELAP standards, where applicable, and except as noted in the laboratory case narrative provided. For a specific list of NELAP and DoD-ELAP-accredited analytes, refer to the certifications section at www.alsglobal.com. Results are intended to be considered in their entirety and apply only to the samples analyzed and reported herein.

If you have any questions, please call me at (805) 526-7161.

Respectfully submitted,

ALS | Environmental

By Kelly Horiochi at 2:41 pm, Jan 16, 2016

For Sue Anderson
Project Manager



2655 Park Center Dr., Suite A
Simi Valley, CA 93065
T: +1 805 526 7161
F: +1 805 526 7270
www.alsglobal.com

Client: Southern California Gas Company Service Request No: P1600187
Project: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

CASE NARRATIVE

The samples were received intact under chain of custody on January 15, 2016 and were stored in accordance with the analytical method requirements. Please refer to the sample acceptance check form for additional information. The results reported herein are applicable only to the condition of the samples at the time of sample receipt.

Methane Analysis

The Silonite canister samples were analyzed per modified EPA Method TO-3 for methane using a gas chromatograph equipped with a flame ionization detector (FID). This procedure is described in laboratory SOP VOA-TO3C1C6. This method is included on the laboratory's DoD-ELAP scope of accreditation, however it is not part of the NELAP or AIHA-LAP accreditation.

Sulfur Analysis

The Zefon bag samples were analyzed for ten sulfur compounds per ASTM D 5504-12 using a gas chromatograph equipped with a sulfur chemiluminescence detector (SCD). All compounds with the exception of hydrogen sulfide and carbonyl sulfide are quantitated against the initial calibration curve for methyl mercaptan. This method is included on the laboratory's NELAP scope of accreditation, however it is not part of the DoD-ELAP or AIHA-LAP accreditation.

Volatile Organic Compound Analysis

The Silonite canister samples were also analyzed for selected volatile organic compounds in accordance with EPA Method TO-15 from the Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air, Second Edition (EPA/625/R-96/010b), January, 1999. This procedure is described in laboratory SOP VOA-TO15. The analytical system was comprised of a gas chromatograph / mass spectrometer (GC/MS) interfaced to a whole-air preconcentrator. This method is included on the laboratory's NELAP and DoD-ELAP scope of accreditation, however it is not part of the AIHA-LAP accreditation. Any analytes flagged with an X are not included on the NELAP or DoD-ELAP accreditation.

The canisters were cleaned, prior to sampling, down to the method reporting limit (MRL) reported for this project. Please note, projects which require reporting below the MRL could have results between the MRL and method detection limit (MDL) that are biased high.

The results of analyses are given in the attached laboratory report. All results are intended to be considered in their entirety, and ALS Environmental (ALS) is not responsible for utilization of less than the complete report.

Use of ALS Environmental (ALS)'s Name. Client shall not use ALS's name or trademark in any marketing or reporting materials, press releases or in any other manner ("Materials") whatsoever and shall not attribute to ALS any test result, tolerance or specification derived from ALS's data ("Attribution") without ALS's prior written consent, which may be withheld by ALS for any reason in its sole discretion. To request ALS's consent, Client shall provide copies of the proposed Materials or Attribution and describe in writing Client's proposed use of such Materials or Attribution. If ALS has not provided written approval of the Materials or Attribution within ten (10) days of receipt from Client, Client's request to use ALS's name or trademark in any Materials or Attribution shall be deemed denied. ALS may, in its discretion, reasonably charge Client for its time in reviewing Materials or Attribution requests. Client acknowledges and agrees that the unauthorized use of ALS's name or trademark may cause ALS to incur irreparable harm for which the recovery of money damages will be inadequate. Accordingly, Client acknowledges and agrees that a violation shall justify preliminary injunctive relief. For questions contact the laboratory.



2655 Park Center Dr., Suite A
 Simi Valley, CA 93065
 T: +1 805 526 7161
 F: +1 805 526 7270
www.alsglobal.com

ALS Environmental – Simi Valley

CERTIFICATIONS, ACCREDITATIONS, AND REGISTRATIONS

Agency	Web Site	Number
AIHA	http://www.aihaaccreditedlabs.org	101661
Arizona DHS	http://www.azdhs.gov/lab/license/env.htm	AZ0694
DoD ELAP	http://www.pjlabs.com/search-accredited-labs	L15-398
Florida DOH (NELAP)	http://www.doh.state.fl.us/lab/EnvLabCert/WaterCert.htm	E871020
Maine DHHS	http://www.maine.gov/dhhs/mecdc/environmental-health/water/dwp-services/labcert/labcert.htm	2014025
Minnesota DOH (NELAP)	http://www.health.state.mn.us/accreditation	977273
New Jersey DEP (NELAP)	http://www.nj.gov/dep/oqa/	CA009
New York DOH (NELAP)	http://www.wadsworth.org/labcert/elap/elap.html	11221
Oregon PHD (NELAP)	http://public.health.oregon.gov/LaboratoryServices/EnvironmentalLaboratoryAccreditation/Pages/index.aspx	4068-001
Pennsylvania DEP	http://www.depweb.state.pa.us/labs	68-03307 (Registration)
Texas CEQ (NELAP)	http://www.tceq.texas.gov/field/qa/env_lab_accreditation.html	T104704413-15-6
Utah DOH (NELAP)	http://www.health.utah.gov/lab/labimp/certification/index.html	CA01627201 5-5
Washington DOE	http://www.ecy.wa.gov/programs/eap/labs/lab-accreditation.html	C946

Analyses were performed according to our laboratory's NELAP and DoD-ELAP approved quality assurance program. A complete listing of specific NELAP and DoD-ELAP certified analytes can be found in the certifications section at www.alsglobal.com, or at the accreditation body's website.

Each of the certifications listed above have an explicit Scope of Accreditation that applies to specific matrices/methods/analytes; therefore, please contact the laboratory for information corresponding to a particular certification.

ALS ENVIRONMENTAL

DETAIL SUMMARY REPORT

Client: Southern California Gas Company
 Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

Service Request: P1600187

Date Received: 1/15/2016
 Time Received: 09:20

Client Sample ID	Lab Code	Matrix	Date Collected	Time Collected	Container ID	Pi1 (psig)	Pf1 (psig)	TO-3 Modified - Cl C6+ Can		
								TO-15 - VOC Cans	ASTM D 5504-12 - Sulfur Bag	
Mason/Corbin	P1600187-001	Air	1/15/2016	02:25	AS00950	-1.66	1.13	X	X	X
Porter Ranch Community School	P1600187-002	Air	1/15/2016	02:50	AS00953	-1.10	1.01	X	X	X
Galileo/Donatello	P1600187-003	Air	1/15/2016	03:14	AS00975	-2.22	1.03	X	X	X
Porter Ranch Estates	P1600187-004	Air	1/15/2016	03:37	AS00958	-1.98	1.02	X	X	X
Highlands 1	P1600187-005	Air	1/15/2016	03:56	AS00916	-1.30	1.00	X	X	X
Highlands 3	P1600187-006	Air	1/15/2016	04:15	AS00959	-2.66	1.00	X	X	X
Porter Ranch Estates 3	P1600187-007	Air	1/15/2016	04:37	AS00930	-1.95	1.11	X	X	X
Highlands 2	P1600187-008	Air	1/15/2016	04:56	AS00985	-1.63	1.13	X	X	X
Castlebay Elementary School	P1600187-009	Air	1/15/2016	05:13	AS00960	-1.56	1.09	X	X	X
Starter Set Preschool	P1600187-010	Air	1/15/2016	05:37	AS00988	-1.86	1.06	X	X	X
Porter Ridge Park	P1600187-011	Air	1/15/2016	07:39	AS00997	-1.75	1.04	X	X	X



Air - Chain of Custody Record & Analytical Service Request

2655 Park Center Drive, Suite A
 Simi Valley, California 93065
 Phone (805) 526-7161
 Fax (805) 526-7270

Requested Turnaround Time in Business Days (Surcharges) please circle
 1 Day (100%) 2 Day (75%) 3 Day (50%) 4 Day (35%) 5 Day (25%) 10 Day-Standard

ALS Project No. 21600187

ALS Contact: Sue Anderson

Company Name & Address (Reporting Information)		Project Name	
AIRKINETICS, INC. 1308 S. Allec Street Anaheim, CA 92805		SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION	
Project Manager SON BUI		Project Number 14424	
Phone (714) 254-1945		P.O. # / Billing Information	
Fax (714) 956-2350		Sampler (Print & Sign)	
Email Address for Result Reporting		Sue Anderson	

Client Sample ID	Laboratory ID Number	Date Collected	Time Collected	Collection Vessel	Canister ID (Bar code # - AC, SC, etc.)	Flow Controller ID (Bar code # - FC #)	Canister Start Pressure "Hg	Canister End Pressure "Hg/psig	Analysis Method		Comment e.g. Actual Preservative or specific instructions
									TO-3 modified for Methane	ASTM D 5504-12 (Selected sulfur compounds & TRS as H2S)	
Mason/Corbin	1	01/15/16	0215 0225	Silonite Can.	A500950	0A00503	29	4	X	TO-15 (BTEX)	
Porter Ranch Community School	2	01/15/16	0240 0250	Tedlar Bag	NA	NA	NA	NA	X		
Galileo/Donatello	3	01/15/16	0304 0314	Silonite Can.	A500955	0A01056	29	4	X		
Porter Ranch Estates	4	01/15/16	0327 0337	Tedlar Bag	NA	NA	NA	NA	X		
Highlands 1	5	01/15/16	0346 0356	Silonite Can.	A500975	0A01371	25	3	X		
Highlands 3	6	0405 01/15/16	0427 0437	Tedlar Bag	NA	NA	NA	NA	X		
Porter Ranch Estates 3	7	01/15/16	0427 0437	Silonite Can.	A500978	0A01867	28.83	7	X		
				Tedlar Bag	NA	NA	NA	NA	X		
				Silonite Can.	A500916	0A00525	27	4	X		
				Tedlar Bag	NA	NA	NA	NA	X		
				Silonite Can.	A500979	0A01268	29.70	25.5-24	X		
				Tedlar Bag	NA	NA	NA	NA	X		
				Silonite Can.	A500959	0A00100	NA	NA	X		
				Tedlar Bag	NA	NA	NA	NA	X		
				Silonite Can.	A500930	0A01268	30	5	X		
				Tedlar Bag	NA	NA	NA	NA	X		

Report Tier Levels - please select

Tier I - Results (Default if not specified) _____ EDD required Yes No Units: _____

Tier II (Results + QC Summaries) X _____

Tier III (Results + QC & Calibration Summaries) _____

Tier IV (Data Validation Package) 10% Surcharge _____

Chain of Custody Seal: (Circle) INTACT BROKEN ABSENT

Relinquished by: (Signature) <u>R. Sabarwal</u>	Received by: (Signature) _____
Date: <u>01/16/16</u>	Date: <u>1/15/16</u>
Time: <u>0920</u>	Time: <u>0910</u>
Relinquished by: (Signature) _____	Received by: (Signature) _____
Date: _____	Date: _____
Time: _____	Time: _____



Air - Chain of Custody Record & Analytical Service Request

2655 Park Center Drive, Suite A
 Simi Valley, California 93065
 Phone (805) 526-7161
 Fax (805) 526-7270

Requested Turnaround Time in Business Days (Surcharges) please circle
 1 Day (100%) 2 Day (75%) 3 Day (50%) 4 Day (35%) 5 Day (25%) 10 Day-Standard

ALS Project No. 911000187

Company Name & Address (Reporting Information)		Project Name	
AIRKINETICS, INC. 1308 S. Allec Street Anaheim, CA 92805		SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION	
Project Manager SON BUJ		Project Number 14424	
Phone (714) 254-1945		P.O. # / Billing Information	
Fax (714) 956-2350		ALS Contact: Sue Anderson	
Email Address for Result Reporting		Analysis Method	
Please see Kelly Horiuchi for distribution list.		TO-3 modified for Methane	

Client Sample ID	Laboratory ID Number	Date Collected	Time Collected	Collection Vessel	Canister ID (Bar code # - AC, SC, etc.)	Flow Controller ID (Bar code # - FC #)	Canister Start Pressure "Hg	Canister End Pressure "Hg/psig	Sampler (Print & Sign)		Comment e.g. Actual Preservative or specific instructions
									Canister	Canister	
Highlands 2	8	01/15/16	0446 0456	Silonite Can. Tedlar Bag	A500985 NA	0A01483 NA	2.8 NA	3.5 NA	X	X	ASTM D 5504-12 (Selected sulfur compounds & TRS as H2S)
Castlebay Elementary School	9	01/15/16	0503 0513	Silonite Can. Tedlar Bag	A500960 NA	0A01237 NA	2.6 NA	2 NA	X	X	
Starter Set Preschool	10	01/15/16	0527 0537	Silonite Can. Tedlar Bag	A500988 NA	0A00704 NA	27.5 NA	4 NA	X	X	
Porter Ridge Park	11	01/15/16	0729 0739	Silonite Can. Tedlar Bag	A500997 NA	0A01542 NA	29.5 NA	4.5 NA	X	X	

Report Tier Levels - please select

Tier I - Results (Default if not specified) _____ EDD required Yes / No

Tier II (Results + QC Summaries) X

Tier III (Results + QC & Calibration Summaries) _____

Tier IV (Data Validation Package) 10% Surcharge Type: _____ Units: _____

Chain of Custody Seal: (Circle) INTACT BROKEN ABSENT

Relinquished by: (Signature) <u>R. Subramanyam</u>	Date: <u>01/15/16</u>	Time: <u>0920</u>	Received by: (Signature) <u>[Signature]</u>	Date: <u>1/15/16</u>	Time: <u>0920</u>
Relinquished by: (Signature)	Date:	Time:	Received by: (Signature)	Date:	Time:

**ALS Environmental
Sample Acceptance Check Form**

Client: Southern California Gas Company

Work order: P1600187

Project: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

Sample(s) received on: 1/15/16

Date opened: 1/15/16

by: KKELPE

Note: This form is used for all samples received by ALS. The use of this form for custody seals is strictly meant to indicate presence/absence and not as an indication of compliance or nonconformity. Thermal preservation and pH will only be evaluated either at the request of the client and/or as required by the method/SOP.

- | | Yes | No | N/A |
|---|-------------------------------------|-------------------------------------|-------------------------------------|
| 1 Were sample containers properly marked with client sample ID? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 2 Did sample containers arrive in good condition? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3 Were chain-of-custody papers used and filled out? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 4 Did sample container labels and/or tags agree with custody papers? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5 Was sample volume received adequate for analysis? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 6 Are samples within specified holding times? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 7 Was proper temperature (thermal preservation) of cooler at receipt adhered to? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 8 Were custody seals on outside of cooler/Box/Container? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| Location of seal(s)? _____ Sealing Lid? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Were signature and date included? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Were seals intact? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 9 Do containers have appropriate preservation , according to method/SOP or Client specified information? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Is there a client indication that the submitted samples are pH preserved? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Were VOA vials checked for presence/absence of air bubbles? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Does the client/method/SOP require that the analyst check the sample pH and <u>if necessary</u> alter it? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 10 Tubes: Are the tubes capped and intact? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 11 Badges: Are the badges properly capped and intact? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Are dual bed badges separated and individually capped and intact? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |

Lab Sample ID	Container Description	Required pH *	Received pH	Adjusted pH	VOA Headspace (Presence/Absence)	Receipt / Preservation Comments
P1600187-001.01	6.0 L Silonite Can					
P1600187-001.02	1 L Zefon Bag					
P1600187-002.01	6.0 L Silonite Can					
P1600187-002.02	1 L Zefon Bag					
P1600187-003.01	6.0 L Silonite Can					
P1600187-003.02	1 L Zefon Bag					
P1600187-004.01	6.0 L Silonite Can					
P1600187-004.02	1 L Zefon Bag					
P1600187-005.01	6.0 L Silonite Can					
P1600187-005.02	1 L Zefon Bag					
P1600187-006.01	6.0 L Silonite Can					
P1600187-006.02	1 L Zefon Bag					
P1600187-007.01	6.0 L Silonite Can					
P1600187-007.02	1 L Zefon Bag					
P1600187-008.01	6.0 L Silonite Can					

Explain any discrepancies: (include lab sample ID numbers): _____

RSK - MEEPP, HCL (pH<2); RSK - CO₂, (pH 5-8); Sulfur (pH>4)

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

ALS Project ID: P1600187

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

Methane

Test Code: EPA TO-3 Modified
 Instrument ID: HP5890 II/GC8/FID
 Analyst: Mike Conejo
 Sampling Media: 6.0 L Silonite Canister(s)
 Test Notes:

Date(s) Collected: 1/15/16
 Date Received: 1/15/16
 Date Analyzed: 1/15/16

Client Sample ID	ALS Sample ID	Canister Dilution Factor	Injection Volume ml(s)	Result ppmV	MRL ppmV	Data Qualifier
Mason/Corbin	P1600187-001	1.21	1.0	4.4	0.61	
Porter Ranch Community School	P1600187-002	1.16	1.0	3.1	0.58	
Galileo/Donatello	P1600187-003	1.26	1.0	3.3	0.63	
Porter Ranch Estates	P1600187-004	1.24	1.0	3.1	0.62	
Highlands 1	P1600187-005	1.17	1.0	2.8	0.59	
Highlands 3	P1600187-006	1.30	1.0	3.0	0.65	
Porter Ranch Estates 3	P1600187-007	1.24	1.0	3.0	0.62	
Highlands 2	P1600187-008	1.21	1.0	2.9	0.61	
Castlebay Elementary School	P1600187-009	1.20	1.0	3.0	0.60	
Starter Set Preschool	P1600187-010	1.23	1.0	2.9	0.62	
Porter Ridge Park	P1600187-011	1.22	1.0	3.0	0.61	
Method Blank	P160115-MB	1.00	1.0	ND	0.50	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

LABORATORY CONTROL SAMPLE SUMMARY

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Lab Control Sample

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187

ALS Sample ID: P160115-LCS

Test Code: EPA TO-3 Modified

Instrument ID: HP5890 II/GC8/FID

Analyst: Mike Conejo

Sampling Media: 6.0 L Silonite Canister

Test Notes:

Date Collected: NA

Date Received: NA

Date Analyzed: 1/15/16

Volume(s) Analyzed: NA ml(s)

Compound	Spike Amount ppmV	Result ppmV	% Recovery	ALS	Data Qualifier
				Acceptance Limits	
Methane	1,020	969	95	83-107	

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company
Client Sample ID: Mason/Corbin
Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187
 ALS Sample ID: P1600187-001

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 7890A/GC22/SCD
 Analyst: Mike Conejo
 Sample Type: 1 L Zefon Bag
 Test Notes:

Date Collected: 1/15/16
 Time Collected: 02:25
 Date Received: 1/15/16
 Date Analyzed: 1/15/16
 Time Analyzed: 09:57
 Volume(s) Analyzed: 2.0 ml(s)

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	5.0	
74-93-1	Methyl Mercaptan	ND	2.5	
75-08-1	Ethyl Mercaptan	ND	2.5	
75-18-3	Dimethyl Sulfide	ND	2.5	
75-15-0	Carbon Disulfide	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	2.5	
75-66-1	tert-Butyl Mercaptan	ND	2.5	
107-03-9	n-Propyl Mercaptan	ND	2.5	
110-01-0	Tetrahydrothiophene	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company
Client Sample ID: Porter Ranch Community School
Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187
 ALS Sample ID: P1600187-002

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 7890A/GC22/SCD
 Analyst: Mike Conejo
 Sample Type: 1 L Zefon Bag
 Test Notes:

Date Collected: 1/15/16
 Time Collected: 02:50
 Date Received: 1/15/16
 Date Analyzed: 1/15/16
 Time Analyzed: 10:13
 Volume(s) Analyzed: 2.0 ml(s)

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	5.0	
74-93-1	Methyl Mercaptan	ND	2.5	
75-08-1	Ethyl Mercaptan	ND	2.5	
75-18-3	Dimethyl Sulfide	ND	2.5	
75-15-0	Carbon Disulfide	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	2.5	
75-66-1	tert-Butyl Mercaptan	ND	2.5	
107-03-9	n-Propyl Mercaptan	ND	2.5	
110-01-0	Tetrahydrothiophene	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company
Client Sample ID: Galileo/Donatello
Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187
 ALS Sample ID: P1600187-003

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 7890A/GC22/SCD
 Analyst: Mike Conejo
 Sample Type: 1 L Zefon Bag
 Test Notes:

Date Collected: 1/15/16
 Time Collected: 03:14
 Date Received: 1/15/16
 Date Analyzed: 1/15/16
 Time Analyzed: 10:25
 Volume(s) Analyzed: 2.0 ml(s)

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	5.0	
74-93-1	Methyl Mercaptan	ND	2.5	
75-08-1	Ethyl Mercaptan	ND	2.5	
75-18-3	Dimethyl Sulfide	ND	2.5	
75-15-0	Carbon Disulfide	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	2.5	
75-66-1	tert-Butyl Mercaptan	ND	2.5	
107-03-9	n-Propyl Mercaptan	ND	2.5	
110-01-0	Tetrahydrothiophene	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company
Client Sample ID: Porter Ranch Estates
Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187
 ALS Sample ID: P1600187-004

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 7890A/GC22/SCD
 Analyst: Mike Conejo
 Sample Type: 1 L Zefon Bag
 Test Notes:

Date Collected: 1/15/16
 Time Collected: 03:37
 Date Received: 1/15/16
 Date Analyzed: 1/15/16
 Time Analyzed: 12:45
 Volume(s) Analyzed: 2.0 ml(s)

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	5.0	
74-93-1	Methyl Mercaptan	ND	2.5	
75-08-1	Ethyl Mercaptan	ND	2.5	
75-18-3	Dimethyl Sulfide	ND	2.5	
75-15-0	Carbon Disulfide	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	2.5	
75-66-1	tert-Butyl Mercaptan	ND	2.5	
107-03-9	n-Propyl Mercaptan	ND	2.5	
110-01-0	Tetrahydrothiophene	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company
Client Sample ID: Highlands 1
Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187
 ALS Sample ID: P1600187-005

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 7890A/GC22/SCD
 Analyst: Mike Conejo
 Sample Type: 1 L Zefon Bag
 Test Notes:

Date Collected: 1/15/16
 Time Collected: 03:56
 Date Received: 1/15/16
 Date Analyzed: 1/15/16
 Time Analyzed: 13:43
 Volume(s) Analyzed: 2.0 ml(s)

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	5.0	
74-93-1	Methyl Mercaptan	ND	2.5	
75-08-1	Ethyl Mercaptan	ND	2.5	
75-18-3	Dimethyl Sulfide	ND	2.5	
75-15-0	Carbon Disulfide	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	2.5	
75-66-1	tert-Butyl Mercaptan	ND	2.5	
107-03-9	n-Propyl Mercaptan	ND	2.5	
110-01-0	Tetrahydrothiophene	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company
Client Sample ID: Highlands 3
Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187
 ALS Sample ID: P1600187-006

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 7890A/GC22/SCD
 Analyst: Mike Conejo
 Sample Type: 1 L Zefon Bag
 Test Notes:

Date Collected: 1/15/16
 Time Collected: 04:15
 Date Received: 1/15/16
 Date Analyzed: 1/15/16
 Time Analyzed: 13:59
 Volume(s) Analyzed: 2.0 ml(s)

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	5.0	
74-93-1	Methyl Mercaptan	ND	2.5	
75-08-1	Ethyl Mercaptan	ND	2.5	
75-18-3	Dimethyl Sulfide	ND	2.5	
75-15-0	Carbon Disulfide	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	2.5	
75-66-1	tert-Butyl Mercaptan	ND	2.5	
107-03-9	n-Propyl Mercaptan	ND	2.5	
110-01-0	Tetrahydrothiophene	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company
Client Sample ID: Porter Ranch Estates 3
Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187
 ALS Sample ID: P1600187-007

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 7890A/GC22/SCD
 Analyst: Mike Conejo
 Sample Type: 1 L Zefon Bag
 Test Notes:

Date Collected: 1/15/16
 Time Collected: 04:37
 Date Received: 1/15/16
 Date Analyzed: 1/15/16
 Time Analyzed: 14:18
 Volume(s) Analyzed: 2.0 ml(s)

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	5.0	
74-93-1	Methyl Mercaptan	ND	2.5	
75-08-1	Ethyl Mercaptan	ND	2.5	
75-18-3	Dimethyl Sulfide	ND	2.5	
75-15-0	Carbon Disulfide	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	2.5	
75-66-1	tert-Butyl Mercaptan	ND	2.5	
107-03-9	n-Propyl Mercaptan	ND	2.5	
110-01-0	Tetrahydrothiophene	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company
Client Sample ID: Highlands 2
Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187
 ALS Sample ID: P1600187-008

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 6890A/GC13/SCD
 Analyst: Mike Conejo
 Sample Type: 1 L Zefon Bag
 Test Notes:

Date Collected: 1/15/16
 Time Collected: 04:56
 Date Received: 1/15/16
 Date Analyzed: 1/15/16
 Time Analyzed: 12:46
 Volume(s) Analyzed: 1.0 ml(s)

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	5.0	
74-93-1	Methyl Mercaptan	ND	2.5	
75-08-1	Ethyl Mercaptan	ND	2.5	
75-18-3	Dimethyl Sulfide	ND	2.5	
75-15-0	Carbon Disulfide	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	2.5	
75-66-1	tert-Butyl Mercaptan	ND	2.5	
107-03-9	n-Propyl Mercaptan	ND	2.5	
110-01-0	Tetrahydrothiophene	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company
Client Sample ID: Castlebay Elementary School
Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187
 ALS Sample ID: P1600187-009

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 6890A/GC13/SCD
 Analyst: Mike Conejo
 Sample Type: 1 L Zefon Bag
 Test Notes:

Date Collected: 1/15/16
 Time Collected: 05:13
 Date Received: 1/15/16
 Date Analyzed: 1/15/16
 Time Analyzed: 12:35
 Volume(s) Analyzed: 1.0 ml(s)

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	5.0	
74-93-1	Methyl Mercaptan	ND	2.5	
75-08-1	Ethyl Mercaptan	ND	2.5	
75-18-3	Dimethyl Sulfide	ND	2.5	
75-15-0	Carbon Disulfide	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	2.5	
75-66-1	tert-Butyl Mercaptan	ND	2.5	
107-03-9	n-Propyl Mercaptan	ND	2.5	
110-01-0	Tetrahydrothiophene	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company
Client Sample ID: Starter Set Preschool
Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187
 ALS Sample ID: P1600187-010

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 6890A/GC13/SCD
 Analyst: Mike Conejo
 Sample Type: 1 L Zefon Bag
 Test Notes:

Date Collected: 1/15/16
 Time Collected: 05:37
 Date Received: 1/15/16
 Date Analyzed: 1/15/16
 Time Analyzed: 10:20
 Volume(s) Analyzed: 1.0 ml(s)

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	5.0	
74-93-1	Methyl Mercaptan	ND	2.5	
75-08-1	Ethyl Mercaptan	ND	2.5	
75-18-3	Dimethyl Sulfide	ND	2.5	
75-15-0	Carbon Disulfide	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	2.5	
75-66-1	tert-Butyl Mercaptan	ND	2.5	
107-03-9	n-Propyl Mercaptan	ND	2.5	
110-01-0	Tetrahydrothiophene	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company
Client Sample ID: Porter Ridge Park
Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187
 ALS Sample ID: P1600187-011

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 6890A/GC13/SCD
 Analyst: Mike Conejo
 Sample Type: 1 L Zefon Bag
 Test Notes:

Date Collected: 1/15/16
 Time Collected: 07:39
 Date Received: 1/15/16
 Date Analyzed: 1/15/16
 Time Analyzed: 10:07
 Volume(s) Analyzed: 1.0 ml(s)

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	5.0	
74-93-1	Methyl Mercaptan	ND	2.5	
75-08-1	Ethyl Mercaptan	ND	2.5	
75-18-3	Dimethyl Sulfide	ND	2.5	
75-15-0	Carbon Disulfide	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	2.5	
75-66-1	tert-Butyl Mercaptan	ND	2.5	
107-03-9	n-Propyl Mercaptan	ND	2.5	
110-01-0	Tetrahydrothiophene	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Method Blank

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187

ALS Sample ID: P160115-MB

Test Code: ASTM D 5504-12

Instrument ID: Agilent 6890A/GC13/SCD

Analyst: Mike Conejo

Sample Type: 1 L Zefon Bag

Test Notes:

Date Collected: NA

Time Collected: NA

Date Received: NA

Date Analyzed: 1/15/16

Time Analyzed: 07:01

Volume(s) Analyzed: 1.0 ml(s)

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	5.0	
74-93-1	Methyl Mercaptan	ND	2.5	
75-08-1	Ethyl Mercaptan	ND	2.5	
75-18-3	Dimethyl Sulfide	ND	2.5	
75-15-0	Carbon Disulfide	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	2.5	
75-66-1	tert-Butyl Mercaptan	ND	2.5	
107-03-9	n-Propyl Mercaptan	ND	2.5	
110-01-0	Tetrahydrothiophene	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Method Blank

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187

ALS Sample ID: P160115-MB

Test Code: ASTM D 5504-12

Instrument ID: Agilent 7890A/GC22/SCD

Analyst: Mike Conejo

Sample Type: 1 L Zefon Bag

Test Notes:

Date Collected: NA

Time Collected: NA

Date Received: NA

Date Analyzed: 1/15/16

Time Analyzed: 06:52

Volume(s) Analyzed: 2.0 ml(s)

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	5.0	
74-93-1	Methyl Mercaptan	ND	2.5	
75-08-1	Ethyl Mercaptan	ND	2.5	
75-18-3	Dimethyl Sulfide	ND	2.5	
75-15-0	Carbon Disulfide	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	2.5	
75-66-1	tert-Butyl Mercaptan	ND	2.5	
107-03-9	n-Propyl Mercaptan	ND	2.5	
110-01-0	Tetrahydrothiophene	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

LABORATORY CONTROL SAMPLE SUMMARY

Page 1 of 1

Client: Southern California Gas Company
Client Sample ID: Lab Control Sample
Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187
 ALS Sample ID: P160115-LCS

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 7890A/GC22/SCD
 Analyst: Mike Conejo
 Sample Type: 1 L Zefon Bag
 Test Notes:

Date Collected: NA
 Date Received: NA
 Date Analyzed: 1/15/16
 Volume(s) Analyzed: NA ml(s)

CAS #	Compound	Spike Amount ppbV	Result ppbV	% Recovery	ALS	Data Qualifier
					Acceptance Limits	
7783-06-4	Hydrogen Sulfide	1,000	884	88	65-138	
463-58-1	Carbonyl Sulfide	1,000	823	82	60-135	
74-93-1	Methyl Mercaptan	1,000	834	83	57-140	

ALS ENVIRONMENTAL

LABORATORY CONTROL SAMPLE SUMMARY

Page 1 of 1

Client: Southern California Gas Company
Client Sample ID: Lab Control Sample
Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187
ALS Sample ID: P160115-LCS

Test Code: ASTM D 5504-12
Instrument ID: Agilent 6890A/GC13/SCD
Analyst: Mike Conejo
Sample Type: 1 L Zefon Bag
Test Notes:

Date Collected: NA
Date Received: NA
Date Analyzed: 1/15/16
Volume(s) Analyzed: NA ml(s)

CAS #	Compound	Spike Amount ppbV	Result ppbV	% Recovery	ALS	Data Qualifier
					Acceptance Limits	
7783-06-4	Hydrogen Sulfide	2,000	2,470	124	65-138	
463-58-1	Carbonyl Sulfide	2,000	2,380	119	60-135	
74-93-1	Methyl Mercaptan	2,000	2,370	119	57-140	

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Mason/Corbin

ALS Project ID: P1600187

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Sample ID: P1600187-001

Test Code: EPA TO-15

Date Collected: 1/15/16

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: 1/15/16

Analyst: Simon Cao

Date Analyzed: 1/15/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 1.00 Liter(s)

Test Notes:

Container ID: AS00950

Initial Pressure (psig): -1.66 Final Pressure (psig): 1.13

Canister Dilution Factor: 1.21

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
71-43-2	Benzene	0.26	0.038	
108-88-3	Toluene	0.39	0.16	
100-41-4	Ethylbenzene	ND	0.14	
179601-23-1	m,p-Xylenes	0.18	0.14	
95-47-6	o-Xylene	ND	0.14	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Porter Ranch Community School

ALS Project ID: P1600187

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Sample ID: P1600187-002

Test Code: EPA TO-15

Date Collected: 1/15/16

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: 1/15/16

Analyst: Simon Cao

Date Analyzed: 1/15/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 1.00 Liter(s)

Test Notes:

Container ID: AS00953

Initial Pressure (psig): -1.10 Final Pressure (psig): 1.01

Canister Dilution Factor: 1.16

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
71-43-2	Benzene	0.24	0.036	
108-88-3	Toluene	0.38	0.15	
100-41-4	Ethylbenzene	ND	0.13	
179601-23-1	m,p-Xylenes	0.17	0.13	
95-47-6	o-Xylene	ND	0.13	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Galileo/Donatello

ALS Project ID: P1600187

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Sample ID: P1600187-003

Test Code: EPA TO-15

Date Collected: 1/15/16

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: 1/15/16

Analyst: Simon Cao

Date Analyzed: 1/15/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 1.00 Liter(s)

Test Notes:

Container ID: AS00975

Initial Pressure (psig): -2.22 Final Pressure (psig): 1.03

Canister Dilution Factor: 1.26

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
71-43-2	Benzene	0.24	0.039	
108-88-3	Toluene	0.36	0.17	
100-41-4	Ethylbenzene	ND	0.15	
179601-23-1	m,p-Xylenes	ND	0.15	
95-47-6	o-Xylene	ND	0.15	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Porter Ranch Estates

ALS Project ID: P1600187

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Sample ID: P1600187-004

Test Code: EPA TO-15

Date Collected: 1/15/16

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: 1/15/16

Analyst: Simon Cao

Date Analyzed: 1/15/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 1.00 Liter(s)

Test Notes:

Container ID: AS00958

Initial Pressure (psig): -1.98 Final Pressure (psig): 1.02

Canister Dilution Factor: 1.24

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
71-43-2	Benzene	0.18	0.039	
108-88-3	Toluene	0.24	0.16	
100-41-4	Ethylbenzene	ND	0.14	
179601-23-1	m,p-Xylenes	ND	0.14	
95-47-6	o-Xylene	ND	0.14	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Highlands 1

ALS Project ID: P1600187

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Sample ID: P1600187-005

Test Code: EPA TO-15

Date Collected: 1/15/16

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: 1/15/16

Analyst: Simon Cao

Date Analyzed: 1/15/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 1.00 Liter(s)

Test Notes:

Container ID: AS00916

Initial Pressure (psig): -1.30 Final Pressure (psig): 1.00

Canister Dilution Factor: 1.17

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
71-43-2	Benzene	0.18	0.037	
108-88-3	Toluene	0.23	0.16	
100-41-4	Ethylbenzene	ND	0.13	
179601-23-1	m,p-Xylenes	ND	0.13	
95-47-6	o-Xylene	ND	0.13	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Highlands 3

ALS Project ID: P1600187

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Sample ID: P1600187-006

Test Code: EPA TO-15

Date Collected: 1/15/16

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: 1/15/16

Analyst: Simon Cao

Date Analyzed: 1/15/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 1.00 Liter(s)

Test Notes:

Container ID: AS00959

Initial Pressure (psig): -2.66 Final Pressure (psig): 1.00

Canister Dilution Factor: 1.30

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
71-43-2	Benzene	0.18	0.041	
108-88-3	Toluene	0.23	0.17	
100-41-4	Ethylbenzene	ND	0.15	
179601-23-1	m,p-Xylenes	ND	0.15	
95-47-6	o-Xylene	ND	0.15	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Porter Ranch Estates 3

ALS Project ID: P1600187

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Sample ID: P1600187-007

Test Code: EPA TO-15

Date Collected: 1/15/16

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: 1/15/16

Analyst: Simon Cao

Date Analyzed: 1/15/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 1.00 Liter(s)

Test Notes:

Container ID: AS00930

Initial Pressure (psig): -1.95 Final Pressure (psig): 1.11

Canister Dilution Factor: 1.24

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
71-43-2	Benzene	0.20	0.039	
108-88-3	Toluene	0.28	0.16	
100-41-4	Ethylbenzene	ND	0.14	
179601-23-1	m,p-Xylenes	ND	0.14	
95-47-6	o-Xylene	ND	0.14	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Highlands 2

ALS Project ID: P1600187

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Sample ID: P1600187-008

Test Code: EPA TO-15

Date Collected: 1/15/16

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: 1/15/16

Analyst: Simon Cao

Date Analyzed: 1/15/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 1.00 Liter(s)

Test Notes:

Container ID: AS00985

Initial Pressure (psig): -1.63 Final Pressure (psig): 1.13

Canister Dilution Factor: 1.21

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
71-43-2	Benzene	0.20	0.038	
108-88-3	Toluene	0.27	0.16	
100-41-4	Ethylbenzene	ND	0.14	
179601-23-1	m,p-Xylenes	ND	0.14	
95-47-6	o-Xylene	ND	0.14	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Castlebay Elementary School

ALS Project ID: P1600187

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Sample ID: P1600187-009

Test Code: EPA TO-15

Date Collected: 1/15/16

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: 1/15/16

Analyst: Simon Cao

Date Analyzed: 1/15/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 1.00 Liter(s)

Test Notes:

Container ID: AS00960

Initial Pressure (psig): -1.56 Final Pressure (psig): 1.09

Canister Dilution Factor: 1.20

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
71-43-2	Benzene	0.21	0.038	
108-88-3	Toluene	0.31	0.16	
100-41-4	Ethylbenzene	ND	0.14	
179601-23-1	m,p-Xylenes	0.14	0.14	
95-47-6	o-Xylene	ND	0.14	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Starter Set Preschool

ALS Project ID: P1600187

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Sample ID: P1600187-010

Test Code: EPA TO-15

Date Collected: 1/15/16

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: 1/15/16

Analyst: Simon Cao

Date Analyzed: 1/15/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 1.00 Liter(s)

Test Notes:

Container ID: AS00988

Initial Pressure (psig): -1.86 Final Pressure (psig): 1.06

Canister Dilution Factor: 1.23

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
71-43-2	Benzene	0.22	0.039	
108-88-3	Toluene	0.32	0.16	
100-41-4	Ethylbenzene	ND	0.14	
179601-23-1	m,p-Xylenes	0.15	0.14	
95-47-6	o-Xylene	ND	0.14	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Porter Ridge Park

ALS Project ID: P1600187

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Sample ID: P1600187-011

Test Code: EPA TO-15

Date Collected: 1/15/16

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: 1/15/16

Analyst: Simon Cao

Date Analyzed: 1/15/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 1.00 Liter(s)

Test Notes:

Container ID: AS00997

Initial Pressure (psig): -1.75 Final Pressure (psig): 1.04

Canister Dilution Factor: 1.22

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
71-43-2	Benzene	0.21	0.038	
108-88-3	Toluene	0.29	0.16	
100-41-4	Ethylbenzene	ND	0.14	
179601-23-1	m,p-Xylenes	ND	0.14	
95-47-6	o-Xylene	ND	0.14	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Method Blank

ALS Project ID: P1600187

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Sample ID: P160115-MB

Test Code: EPA TO-15

Date Collected: NA

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: NA

Analyst: Simon Cao

Date Analyzed: 1/15/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 1.00 Liter(s)

Test Notes:

Canister Dilution Factor: 1.00

CAS #	Compound	Result ppbV	MRL ppbV	Data Qualifier
71-43-2	Benzene	ND	0.031	
108-88-3	Toluene	ND	0.13	
100-41-4	Ethylbenzene	ND	0.12	
179601-23-1	m,p-Xylenes	ND	0.12	
95-47-6	o-Xylene	ND	0.12	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

SURROGATE SPIKE RECOVERY RESULTS

Page 1 of 1

Client: Southern California Gas Company

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187

Test Code: EPA TO-15

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Analyst: Simon Cao

Sample Type: 6.0 L Silonite Canister(s)

Test Notes:

Date(s) Collected: 1/15/16

Date(s) Received: 1/15/16

Date(s) Analyzed: 1/15/16

Client Sample ID	ALS Sample ID	1,2-Dichloroethane-d4	Toluene-d8	Bromofluorobenzene	Acceptance Limits	Data Qualifier
		Percent Recovered	Percent Recovered	Percent Recovered		
Method Blank	P160115-MB	100	100	106	70-130	
Lab Control Sample	P160115-LCS	92	98	109	70-130	
Mason/Corbin	P1600187-001	99	98	111	70-130	
Porter Ranch Community School	P1600187-002	104	96	109	70-130	
Galileo/Donatello	P1600187-003	104	96	109	70-130	
Porter Ranch Estates	P1600187-004	111	95	109	70-130	
Highlands 1	P1600187-005	108	98	104	70-130	
Highlands 3	P1600187-006	106	100	102	70-130	
Porter Ranch Estates 3	P1600187-007	110	100	103	70-130	
Highlands 2	P1600187-008	110	98	102	70-130	
Castlebay Elementary School	P1600187-009	111	99	102	70-130	
Starter Set Preschool	P1600187-010	112	98	100	70-130	
Porter Ridge Park	P1600187-011	113	99	100	70-130	

Surrogate percent recovery is verified and accepted based on the on-column result.

Reported results are shown in concentration units and as a result of the calculation, may vary slightly from the on-column percent recovery.

ALS ENVIRONMENTAL

LABORATORY CONTROL SAMPLE SUMMARY

Page 1 of 1

Client: Southern California Gas Company

Client Sample ID: Lab Control Sample

Client Project ID: SOUTHERN CALIFORNIA GAS - ALISO CANYON STATION / 14424

ALS Project ID: P1600187

ALS Sample ID: P160115-LCS

Test Code: EPA TO-15

Date Collected: NA

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: NA

Analyst: Simon Cao

Date Analyzed: 1/15/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 0.125 Liter(s)

Test Notes:

CAS #	Compound	Spike Amount ppbV	Result ppbV	% Recovery	ALS	Data Qualifier
					Acceptance Limits	
71-43-2	Benzene	70.8	59.2	84	61-110	
108-88-3	Toluene	57.9	48.3	83	67-117	
100-41-4	Ethylbenzene	50.2	43.4	86	69-123	
179601-23-1	m,p-Xylenes	98.6	86.0	87	67-125	
95-47-6	o-Xylene	48.4	41.6	86	67-124	

Laboratory Control Sample percent recovery is verified and accepted based on the on-column result.
Reported results are shown in concentration units and as a result of the calculation, may vary slightly.